

# CERTIFICATE OF CONFORMITY

No.: PTC20042500801E-EM01

**Applicant** :  
**Address** :  
**Manufacturer** :  
**Address** :  
**Trade Mark** :  
**Product** : **Infrared forehead thermometer**  
**Model** : **T2;T1;T9**

The submitted sample of the above equipment has been tested and found to comply with the following European Directive:

**EMC Directive - 2014/30/EU**

The standard(s) used for showing compliance with the essential requirements:

Applicable Standard(s)	Test Report(s) Number
EN 55011:2016+A1:2017 EN 60601-1-2:2017 EN 61000-3-2:2014 EN 61000-3-3:2013	PTC20042500801E-EM01

This certificate is part of the full test report(s) and should be read in conjunction with it. This certificate is based on an evaluation of one sample of above mentioned product. It does not imply assessment of the production of the product. Without the written approval of Precise Testing & Certification (Guangdong) Co., Ltd. this certificate is not permitted to be reproduced, except in full. It is not permitted to use the test lab's logo. The CE marking may only be used if all the relevant and effective European Directives are applicable.



**Jacky Ou**  
**Manager**  
**Date: April 28, 2020**

